



PAMAS 4132 **Online particle counter** **for filter test rigs**

Single Pass and Multipass filter efficiency analysis

- Analysis of high concentration levels in β -ratio applications
- High resolution analysis in 32 size channels
- Sensors can handle high pressures
- Remotely adjustable thresholds
- Different sensors for various applications

Due to the volumetric cell design of PAMAS particle sensors, 100% of the sample flow volume is analysed this guarantees highest volume accuracy and reproducibility and best statistical information:

- 1,5 $\mu\text{m(c)}$ sensitivity according to ISO 11171
- 1 μm sensitivity according to ISO 4402
- 0,5 μm sensitivity according to ISO 21501

PAMAS 4132

Standardised β -ratio analysis



The **PAMAS 4132** is used as a particle counting system for Single & Multipass Filter test rigs. The standard system is designed to fit perfectly in the test rig of all manufacturers. With Multipass filter testing, the oil circulates in the test stand; with Single Pass filter testing, the liquid passes the filter only once and is then deviated to the out-flow.

For β -ratio analyses, the Single Pass Filter Test is performed with two particle counters. Two instruments - equipped with 32 size channels each - are also used for the Multipass Filter Test. The test rig with two particle

counters allows simultaneous upstream and downstream measurements. During the Multipass Filter Test, the liquid continuously flows through the system.

The online particle counting system **PAMAS 4132** analyses liquids of various viscosity degrees (e.g. oil, fuel, water, etc.). The system is adaptable to any filter test application.

Calibration of particle sensor:

The particle sensor is calibrated according to International Calibration Standards which are traceable to the NIST (National Institute of Standards and Technology).

Single particle counting system using volumetric sensor cells

A highly sophisticated sensor cell and optics guarantees best resolution and accuracy even under high pressure conditions. Particle counting can be achieved using many methods, but only the use of volumetric cells, like those used in PAMAS sensors, can guarantee that all particles passing through the sensor are counted. This results in better statistical analysis and prevents the loss of information compared to in-situ cells that detect only a small portion of the whole sample flow, especially as the samples are getting cleaner.



Due to its compact design, the online particle counting system PAMAS 4132 can be easily integrated into existing test rigs.

Technical data

Power supply:

100 - 240 V AC / 50-60 Hz

Data transfer:

via Ethernet

Volumetric particle sensors:

PAMAS HCB-LD-50/50

Calibration range:

- 1-400 μm according to ISO 21501
- 1-100 μm according to ISO 4402
- 4-70 $\mu\text{m(c)}$ according to ISO 11171

Maximum particle concentration:
24,000 p/ml* at a flow rate of
25 ml/min**

PAMAS HCB-LD-25/25

Calibration range:

- 1-200 μm according to ISO 21501
- 1-100 μm according to ISO 4402
- 4-70 $\mu\text{m(c)}$ according to ISO 11171

Maximum particle concentration:
120,000 p/ml* at a flow rate of
25 ml/min**

PAMAS HCB-LD-15/25

Calibration range:

- 1-100 μm according to ISO 21501
- 1-100 μm according to ISO 4402
- 4-70 $\mu\text{m(c)}$ according to ISO 11171

Maximum particle concentration:
200,000 p/ml* at a flow rate of
10 ml/min**

PAMAS SLS-25/25

Calibration range:

- 0.5-20 μm according to ISO 21501
- 1-20 μm according to ISO 4402
- 1.5-25 $\mu\text{m(c)}$ according to ISO 11171

Maximum particle concentration:
13,000 p/ml* at a flow rate of 10 ml/
min**

* Coincidence error of 7.8%.

**Various flow rates are available.

Temperature:

- Fluid temperature: 0-60° C
- Calibration temperature: 20 or 40° C

Size:

200 mm x 300 mm x
330 mm (L x W x H)



Management System
ISO 9001:2015

www.tuv.com
ID 10105038017

PAMAS HEAD OFFICE Dieselstraße 10, D-71277 Rutesheim, Phone: +49 7152 99 63 0, Fax: +49 7152 99 63-32, Email: info@pamas.de
PAMAS USA 1408 South Denver Avenue, Tulsa, OK 74119 USA, Phone: +1 918 743 6762, Fax: +1 918 743 6917, Email: clay.bielo@pamas.de
PAMAS BENELUX Mechelen Campus, Schaliënhoedreef 20T, B-2800 Mechelen, Phone: +32 15 28 20 10, Mobile: +32 477 42 48 62, Email: paul.pollmann@pamas.de
PAMAS FRANCE Route du Tailleur 210/136, F-40170 Saint-Julien-en-Born, Mobile +33 6 25 33 20 41, Email: eric.colon@pamas.fr
PAMAS LATIN AMERICA Curitiba-Paraná, Brazil, Phone/Fax: +55 41 3022 5445, Mobile: +55 41 999 72 21 73, Email: marcelo.aiub@pamas.de
PAMAS INDIA No. 203, I floor, Oxford House, #15 Rustam Bagh Main Road, Bangalore 560017, India, Phone: +91 80 41 15 00 39, Email: info@pamas.in
PAMAS HISPANIA Calle Zubilleta No. 13 1ºB, ES-48991 Algorta, Mobile: +34 67 75 39 699, Email: julian.malaina@pamas.de
PAMAS UK Sci-Tech Daresbury, Keckwick Lane, Daresbury, Cheshire WA4 4FS, Mobile: +44 79 17 71 33 66, Email: graeme.oakes@pamas.de

Please visit our website at www.pamas.de